

# TANZANIA NATIONAL ROADS AGENCY



P.O. Box 11364  
3<sup>rd</sup> Floor  
Airtel House  
Ali Hassan Mwinyi / Kawawa Roads  
Junction  
Dar es Salaam

**Date:** 15<sup>th</sup> October 2018

**Our Ref:** TRD/HQ/GEN/1064/01

**To:** All Prospective Bidders

**RE: SUPPLY OF TRAINING EQUIPMENT FOR EMERGENCY MEDICAL SERVICES (EMS) ALONG DAR ES SALAAM – MOROGORO ROAD SECTION, TENDER NO: AE/001/2017-18/HQ/G/108**

**Sub: Clarification No. 1 to the Tender Documents**

Please refer to the above subject.

Pursuant to Sub-Clause 8.1 of the Instructions to Tenderers (ITT) of the Tender Documents, TANROADS issues the attached **Clarification No. 1** to the Tender Documents.

Eng. Patrick A. L. Mfugale  
**CHIEF EXECUTIVE**

**Encl: As stated**

**THE UNITED REPUBLIC OF TANZANIA**

**MINISTRY OF WORKS, TRANSPORT AND COMMUNICATION**



Good roads for national development

**SUPPLY OF TRAINING EQUIPMENT FOR EMERGENCY MEDICAL SERVICES (EMS)  
ALONG DAR ES SALAAM –MOROGORO ROAD SECTION**

**TENDER NO: AE/001/2017-18/HQ/G/108**

**Clarification No. 1 to the Bidding Documents**

**Date:** 15<sup>th</sup> October, 2018

**Clarification No. 1** to the bidding document is issued in accordance with Instructions to Bidders, Sub-Clause 8.1 of the Bidding documents:

**QUESTION NO. 1**

Item No. 55; ULTRASOUND MACHINE – we need more clarification for 3 probes; which type of probes do tender require

***Response to Question No. 1:***

*The Ultrasound Machine shall have Cardiac probe, Linear (vascular) probe and Curved (abdominal) probe.*

**QUESTION NO. 2**

Item No. 65; CPAP MASKS + TUBING – we need more clarification/ information – is it reusable or single patient use device?

***Response to Question No. 2:***

*The CPAP Masks +Tubing shall be re-usable.*

*AS*